# 取扱説明書

ブラケット

# KRB4

Part No. Z1-901-270, IA003853
Feb. 2008

次の手順で機器にブラケットを取り付け、EIA 規格のラックに組み込みます。

1. 付属のねじを使用してブラケットを図のように取り付けます。付属のねじ（M4 x 10 皿ねじ）以外は使用しないでください。




2. 付属のねじとワッシャを使用して EIA 規格のラックに取り付けます。



⚠注意 ・機器が落下する恐れがあります。

使用するラックに適合したサポートアングルを取り付けて、本体を支持ください。




| No. | 説明 | 数量 |
|---|---|---|
| 1 | Bracket KRB4 | 2 |
| 2 | Flat head screw M4 x 0.7 x 10 | 4 |
| 3 | Oval head screw M5 x 0.8 x 14 | 4 |
| 4 | Finishing washer 5CW | 4 |

菊水電子工業株式会社
本社・技術センター　〒224-0023　横浜市都筑区東山田 1-1-3
TEL：045-593-0200（代）
**http://www.kikusui.co.jp/**

Attach the mounting bracket to the instrument and install it to the EIA standard rack in the following procedure.

1. Attach the mounting bracket to the unit using the provided screws as shown in figure. Do not use other than the provided screws (M4 x 10 flat-head screw).




2. Install the rack mounting unit to the EIA standard rack using the provided screws and washers.



**⚠CAUTION** ・The instrument may fall.

Install the suitable support angles applying to the used rack system to support the instrument.




| No. | DESCRIPTION | QTY |
|---|---|---|
| 1 | Bracket KRB4 | 2 |
| 2 | Flat-head screw M4 x 0.7 x 10 | 4 |
| 3 | Oval head M5 x 0.8 x 14 | 4 |
| 4 | Fnishing washer 5CW | 4 |

.

KIKUSUI ELECTRONICS CORP.
1-1-3, Higashiyamata, Tsuzuki-ku, Yokohama, 224-0023, Japan
TEL : +81-45-593-7570 FAX: +81-45-593-7571
**http://www.kikusui.co.jp/**

Printed in Japan